

Manufacturer: Kayserbetten GmbH & Co. KG  
Rieper Str. 12  
D-29683 Bad Fallingbostel-Dorfmark  
Germany

We hereby declare in sole responsibility that the **medical devices (class I)** of the series:

Timmy I	Timmy I 1.0	Hannah farbig	Olaf farbig	Emma farbig
Timmy II	Timmy II 1.0	Hannah natur	Olaf natur	Emma natur
Timmy junior	Timmy junior 1.0	Hannah farbig 1.0	Olaf junior	Emma farbig 1.0
Timmy vario	Timmy vario 1.0	Hannah natur 1.0	Olaf farbig 1.0	Emma natur 1.0
Timmy deluxe	Timmy deluxe 1.0		Olaf natur 1.0	
Timmy EasyLift	Timmy EasyLift 1.0		Olaf junior 1.0	IDA farbig
Timmy Teleskop	Timmy Teleskop 1.0			IDA natur
				IDA farbig 1.0
				IDA natur 1.0
Lotte	Wickeltisch	Therapieliege		

are designed and produced in conformity with the technical documentation and the council directive

**93/42/ECC-Annex VII Directive of Medical Devices**

and all applicable parts of the standard specification:

- DIN EN 50637:2018**
- DIN EN 60601-1:2013**
- DIN EN 60601-2-52:2016**
- DIN 32623:2009**
- DIN EN 716-1:2017**

When relevant changes are made to the above mentioned medical devices the declaration of conformity becomes invalid according to the guidelines of the resolution **EK-Med 3.9 A4 of ZLG.**

Bad Fallingbostel, 01.05.2019



Peter Kayser / managing director



Torsten Kappenberg / managing director